

Cell Surface Protein Processing from Cell Suspensions with Cell Surface-Compatible Universal Cell Capture Kit

In this workflow, cells are collected and then bound to the Universal Cell Capture Beads; then the RNA and Protein samples are prepared for hybridization.

The Cell Surface Compatible Universal Cell Capture Kit can be used with the nCounter® Vantage 3D™ RNA:Protein Immune Cell Profiling Assay for cell suspensions and the Vantage 3D Protein Immune Cell Profiling Panel for cell suspensions.

This protocol is compatible with intact cell suspensions from cell lines, PBMCs, and other primary human cells. FFPE and fresh frozen tissue are not compatible with this protocol. Following sample preparation, the RNA and protein components are combined in a single hybridization reaction. Contact NanoString Support (support@nanosttring.com) to receive additional assistance with this assay.

Table 1. Materials provided by NanoString

Product	Material	Description	Storage
nCounter Vantage 3D™ Protein (D) (ERCC controls)	Protein TagSet	Panel	At or below -80°C
	Antibody Mix	Oligonucleotide tags bound to antibodies	At or below -80°C
or			
nCounter Vantage 3D Protein (R) (no ERCC controls)	Protein Plus	Panel	At or below -80°C
	Antibody Mix	Oligonucleotide tags bound to antibodies	At or below -80°C
or			
nCounter Vantage 3D RNA:Protein (no ERCC controls)	Protein Plus	Panel	At or below -80°C
	Antibody Mix	Oligonucleotide tags bound to antibodies	At or below -80°C

Product	Material	Description	Storage
Cell Surface-compatible Universal Cell Capture Kit	Universal Cell Capture Beads	Magnetic bead solution	4°C (Do not freeze)
	Buffer W	Blocking and Wash buffer	4°C
	Buffer LH	Lysis buffer	RT

FOR RESEARCH USE ONLY. Not for use in diagnostic procedures

© 2017 NanoString Technologies, Inc. All rights reserved.

NanoString, NanoString Technologies, the NanoString logo, nCounter, and Vantage 3D are trademarks or registered trademarks of NanoString Technologies, Inc., in the United States and/or other countries

Table 2. Required materials and reagents

Item	Manufacturer	Part Numbers
96-well plate magnet separator*	Thermo Fisher Scientific	12027
	Stemcell Technologies	18102
96-well clear polystyrene round-bottom plate	Corning	351177
Pipettes for 10–1,000 μL *	Various	Various
Manual multi-channel pipette for 200 μL *	Rainin	L12-200XLS+
12-strip standard tubes*	Bioexpress	T-3034-1
1.7 mL microcentrifuge tubes*	Various	Various
Hemocytometer*	Various	Various
Trypan Blue*	Various	Various
Human TruStain FcX™**	Biolegend	422301 or 422302
1X phosphate buffered saline (PBS; pH 7.4)*	Thermo Fisher Scientific	10010-023

*Alternative products can be used if they offer similar function and reliability.

**Only required for samples containing human Fc receptor (e.g., PBMCs).


Table 3. Estimated timeline for cell surface protein detection from cell suspensions. Hands-on time may be less than the processing times listed below.

Step	Minimum Time Required
Advance Preparation	~1 hour 10 min then overnight
Sample Collection	30 min
Bind Sample to Universal Cell Capture Beads	45–60 min
RNA Sample Preparation	30 min
Protein Sample Preparation	1 hour 30 min

Advance Preparation

The following procedure is used for performing 12 reactions. Scale the number of wells per the number of reactions in your experiment.

1. Pre-block the Protein Sample Plate by adding 300 μ L of Buffer W to each of the 12 wells in two rows of a 96-well round bottom plate.
2. Pre-block RNA Sample Plate by adding 300 μ L of Buffer W to each of the 12 wells in one row of a second 96-well round bottom plate.
3. Block both plates on a flat surface for at least 1 hour at room temperature or overnight at 4°C.



Sample Collection

NOTES:

- This section is performed in the Protein Sample Plate.
 - 50,000 cells (or 100,000 primary cells such as PBMCs) per sample are recommended for use per reaction.
 - The number of cells can be decreased to a minimum of 20,000 cells (or 50,000 primary cells such as PBMC) if cell numbers are limited but note that using less than the recommended cell number above may result in reduced signal. Using more than the recommended cell number may require increasing the volume of Buffer LH. See [Table 4](#) (RNA Sample Preparation) and [Table 5](#) (Protein Sample Preparation) for guidelines.
 - Total lysate volume (combined RNA and protein) added to the CodeSet (MAN-10060) or TagSet (MAN-10065) hybridization should not exceed 5 μ L.
 - Perform steps where temperature is not specified at room temperature.
1. Determine the concentration of total viable cells in each sample.
 2. For each sample, collect a minimum of 20,000 cells (or 50,000 from primary cell samples) in Buffer W, 1X PBS containing 2% FBS, or warmed (37°C) cell culture medium in a 1.7 mL microcentrifuge tube.
 3. Remove and discard Buffer W from the Protein Sample Plate by either:
 - inverting and flicking the plate once, followed by blotting of the inverted plate on a fresh paper towel or lab wipe; **or**
 - pipetting out the buffer.
 4. Transfer cell samples to each well in the **Top** row (A) of the Protein Sample Plate. Add Buffer W if necessary to bring the final volume of each sample well to 200 μ L.



Bind the Sample to Universal Cell Capture Beads

NOTE: Keep the plate in contact with the magnet at all times during the step to avoid sample/bead loss. Removing residual liquid after flicking/blotting is not necessary. If buffer is removed by the pipetting method, remove as much of the residual buffer as possible to avoid leaving variable amounts of remaining buffer in the wells. Failure to do so may result in poor quality data. The minimum time for magnet pulldown of beads is at least 3 min but 5 min is recommended. Less than 3 min may result in sample/bead loss and reduced signal in this assay.

1. Prior to opening the vial of Universal Cell Capture Beads, ensure no beads are on the cap, and then thoroughly re-suspend the beads by pipetting.
2. Add 9 μL of Universal Cell Capture Beads to each sample well (A) in the Protein Sample Plate. Use a new pipette tip for each sample well.
3. Mix the samples using a multichannel pipette set to half the sample volume (100 μL).
4. Incubate plate on a flat surface for 30 min at 4°C.
5. Immobilize the bead/cell complexes by placing the Protein Sample Plate on a 96-well plate magnet. Leave plate on the magnet for 5 min, undisturbed.
6. Remove and discard the supernatant by either:
 - firmly holding the plate on the magnet and inverting and flicking the plate/magnet only once, followed by blotting of the inverted plate on a fresh paper towel or lab wipe; **or**
 - using a single-channel pipette to carefully remove supernatant from each sample well while holding the plate on the magnet.




NOTE: Keep the plate in contact with the magnet at all times during this step.

7. For primary cells and cell lines expressing human Fc receptors (e.g., cells expressing CD16, CD64, and/or CD32), blocking Fc receptor-mediated antibody binding is necessary to avoid increased background in the assay. **If this does NOT apply, proceed to Step 8.**
 - a. Prepare 1X Fc Receptor Blocking Solution by diluting 65 μL of BioLegend TruStain FcX in 585 μL of Buffer W. (Extra volume is included to account for variation in pipetting.)
 - b. Add 50 μL 1X Fc Receptor Blocking Solution to each sample well and thoroughly re-suspend beads by pipetting gently.
 - c. Incubate for 10 min at room temperature.
 - d. Add 150 μL Buffer W to each sample well and proceed directly to [RNA Sample Preparation \(Optional\)](#).
8. Add 200 μL Buffer W to each sample well, and thoroughly re-suspend beads by pipetting gently.

RNA Sample Preparation (Optional)

1. Remove and discard Buffer W from Row A of the **RNA Sample Plate** by either:
 - inverting and flicking the plate once, followed by blotting of the inverted plate on a fresh paper towel or lab wipe; **or**
 - pipetting out the buffer.

2. Mix samples from the [Bind the Sample to Universal Cell Capture Beads](#) section thoroughly by pipetting gently and transfer 130 μL of each sample from Row A of the Protein Sample Plate to pre-blocked Row A of the RNA Sample Plate.



NOTE: RNA and Protein samples may be processed sequentially or in parallel from this point. If sequential sample processing is preferred, keep the Protein Sample Plate on ice until ready to proceed with [Protein Sample Preparation](#).

3. Immobilize the bead/cell complexes by placing the RNA Sample Plate on a 96-well plate magnet. Leave plate on the magnet for 5 minutes, undisturbed.
4. Remove and discard the supernatant by either:
 - firmly holding the plate on the magnet and inverting and flicking the plate/magnet only once, followed by blotting the inverted plate on a fresh paper towel or lab wipe; **or**
 - using a single-channel pipette to carefully remove supernatant from each sample well while holding the plate on the magnet.

NOTE: Keep the plate in contact with the magnet at all times during this step to avoid sample/bead loss.

5. Without disturbing the bead/cell pellets, use a single-channel pipette to carefully remove remaining residual buffer from each sample well.
6. Add Buffer LH to each sample well.

NOTE: The volume of Buffer LH is dependent upon the initial number of cells in each sample. Refer to [Table 4](#) to determine the volume of Buffer LH to add to each sample well. For example, if starting with 20,000 cells for RNA:Protein, add 6 μL of Buffer LH to the RNA Sample Plate.

7. Pipette thoroughly to lyse cells directly on the beads. Incubate the RNA lysates for 2–3 min at room temperature.

NOTE: Avoid creating bubbles during lysis step by setting the pipette to half the volume (e.g., 3 μL in this example of Buffer LH). Failure to do so may result in a loss of sample.

8. (Optional) If the lysate is very viscous, add an additional volume equivalent to Step 6 (e.g., 6 μL in this example) of Buffer LH.

Table 2. Buffer LH volume based on initial total cells.

Initial Total Cells (cell lines)	Initial Total Cells (primary cells)	Buffer LH for RNA Lysates
20,000	50,000	6 μ L
50,000	100,000	12 μ L

- Place the RNA Sample Plate on a 96-well plate magnet. Leave plate on the magnet for 5 minutes, undisturbed. **Do not discard the supernatant.**

NOTE: The minimum time for magnet pulldown of beads is at least 3 min but 5 min is recommended.

- Without disturbing the bead/cell pellets, carefully collect each RNA lysate/supernatant sample and transfer to a 12-well strip tube using a single-channel pipette.
- Keep the lysates on ice until you are ready to perform the CodeSet (MAN-10060) or TagSet (MAN-10065) hybridization. If not using immediately, samples can be stored at -80°C .

Protein Sample Preparation


NOTE:

- Avoid creating bubbles during wash steps by setting pipettes to half the sample volume.
1. Add 130 μL Buffer W to each sample well in Row A in the **Protein Sample Plate** to bring the total sample volume to 200 μL .
 2. Add 10 μL antibody mix (Ab mix) to each sample. Use a new tip for each sample.
 3. Thoroughly mix the samples by pipetting gently with pipette set to half the sample volume (100 μL) and incubate for 30 minutes at 4°C.
 4. Immobilize the bead/cell complex for 5 minutes on the plate magnet.
 5. Remove and discard the supernatant by either:
 - firmly holding the plate **on the magnet** and inverting and flicking the plate/magnet only once, followed by blotting of the inverted plate on a fresh paper towel or lab wipe; **or**
 - using a single-channel pipette to carefully remove supernatant from each sample well while holding the plate **on the magnet**.




NOTE: Keep the plate in contact with the magnet at all times during this step to avoid sample/bead loss.

6. Perform a total of 4 washes as follows:
 - a. Remove the plate from the magnet and add 200 μL Buffer W to each sample well. Mix gently by pipetting.
 - b. Immobilize the bead/cell complex for 5 min on the plate magnet, followed by removal and disposal of the supernatant.
 - c. Repeat Steps 6a–b for a second wash.
 - d. Add 200 μL Buffer W to each sample. Mix gently by pipetting and **transfer the samples to the empty row (B) of pre-blocked wells**.



NOTE: Do not skip this step. Decreased assay sensitivity may occur if this step is not performed.

- e. Immobilize the bead/cell complex for 5 min on the plate magnet, followed by removal and disposal of the supernatant.
 - f. Repeat Steps 6a–b once more. No additional well transfers are necessary.
7. Without disturbing the bead/cell pellets, use a single-channel pipette to carefully remove remaining residual buffer from each sample well.
 8. Add Buffer LH to each sample well.



The volume of Buffer LH is dependent upon the initial number of cells in each sample. Refer to [Table 5](#) to determine the volume of Buffer LH to add to each sample well. For example, if starting with 20,000 cells for RNA:Protein, add 10 μL of Buffer LH.

-
- Pipette thoroughly to lyse cells directly on the beads. Incubate the protein lysates for 2–3 min at room temperature.

NOTE: Avoid creating bubbles during lysis step (by setting the pipette to half the volume e.g., 5 μL in this example of Buffer LH). Failure to do so may result in a loss of sample.

- (Optional) If the lysate is very viscous, add an additional volume equivalent to Step 6 (e.g., 10 μL in this example) of Buffer LH.

Table 3. Buffer LH volume based on initial total cell lines.

Initial Total Cells (cell lines and primary cells)	Buffer LH for Protein Lysates
20,000	10 μL
50,000	25 μL
100,000	50 μL

- Place the Protein Sample Plate on a 96-well plate magnet to immobilize the bead/cell complexes. Leave plate on the magnet for 5 minutes, undisturbed. **Do not discard the supernatant.**
- Without disturbing the bead/cell pellets, carefully collect each protein lysate/supernatant sample and transfer to a 12-well strip tube using a single-channel pipette.
- Cap the tubes and denature protein lysates only by incubating for 15 min at 95°C in a thermocycler with a heated lid at 100°C, and then immediately ramp down to 4°C or snap cool on ice for a minimum of 2 minutes.
- Keep the lysates on ice until you are ready to perform CodeSet (MAN-10060) or TagSet (MAN-10065) hybridization. If not using immediately, samples can be stored at -80°C.

NOTE: Denaturation of protein lysates is critical for optimal assay performance. It is not necessary to denature the RNA lysates.